



CB TEST CERTIFICATE

Ref. Certificate No.

NL-10213

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

Issued by:	KEMA Quality B.V.		
Registration date:	2006-02-28		
Product:	Gas Analyzer		
Applicant:	Sick Maihak GmbH	Nimburger Strasse 11	Germany
Manufacturer:	Maihak AG	D-79278 Reute Poppenbuetteler Bogen 9B	Germany
Factory:	Maihak AG	22399 Hamburg Poppenbuetteler Bogen 9B 22399 Hamburg	Germany
Rating and principal characteristics:	100 / 115 / 230 Vac, 48 - 62 Hz; 20 - 200 VA		
Trade mark (if any):	SICK MAIHAK		
Model/Type reference:	SIDOR, S710, S711 and S715		
Additional information:	-		
Sample of product tested to be in conformity with IEC:	61010-1(ed.2)		
Test Report Ref. No:	2082921.50-QUA/ESG		

This CB Test Certificate is issued by the National Certification Body:

KEMA Quality B.V.
Utrechtseweg 310
P.O. Box 5185
6802 ED Arnhem
The Netherlands



Signed by: A.G.H. Bergervoet

Date of issue: 2006-02-28