

# PRODUCT CONFORMITY CERTIFICATE

This is to certify that the

**MAC GMS800**

Manufactured by:

**SICK AG**

Poppenbütteler Bogen 9b  
22399 Hamburg  
Germany

Has been assessed by Sira Certification Service  
And for the conditions stated on this certificate complies with:

**MCERTS Performance Standards for Continuous Emission  
Monitoring Systems, Version 3.4 dated July 2012  
EN15267-1:2009, EN15267-2:2009, EN15267-3:2007  
& QAL 1 as defined in EN 14181: 2004**

Certification Ranges :

**CO** : 0 to 75 mg/m<sup>3</sup> (UNOR) and 0 to 200 mg/m<sup>3</sup> (MULTOR);  
**NO** : 0 to 100 mg/m<sup>3</sup> (UNOR) and 0 to 250 mg/m<sup>3</sup> (MULTOR) and 0 to 50 mg/m<sup>3</sup> (DEFOR);  
**NO<sub>2</sub>** : 0 to 50 mg/m<sup>3</sup> (DEFOR); **NO<sub>x</sub>** : 0 to 100 mg/m<sup>3</sup> (UNOR)  
**SO<sub>2</sub>** : 0 to 75 mg/m<sup>3</sup> (UNOR) and 0 to 250 mg/m<sup>3</sup> (MULTOR) and 0 to 75 mg/m<sup>3</sup> (DEFOR);  
**CH<sub>4</sub>** : 0 to 50 mg/m<sup>3</sup> (UNOR) and 0 to 286 mg/m<sup>3</sup> (MULTOR); **N<sub>2</sub>O** : 0 to 50 mg/m<sup>3</sup> (UNOR); **CO<sub>2</sub>** :  
0 to 25 %vol. (UNOR and MULTOR); **O<sub>2</sub>** : 0 to 25 %vol. (OXOR-P and OXOR-E)

*\*\*See description for additional measuring ranges\*\**

Project No. : 16A23053  
Certificate No : Sira MC100182/04  
Initial Certification : 21 December 2010  
This Certificate issued : 21 December 2015  
Renewal Date : 20 December 2020

Joe Prince MSc, MInst MC  
Deputy Certification Manager

MCERTS is operated on behalf of the Environment Agency by

## Sira Certification Service

Unit 6, Hawarden Industrial Park  
Hawarden, Deeside, CH5 3US  
Tel: +44 (0)1244 670 900



**sira**  
CERTIFICATION

*The MCERTS certificate consists of this document in its entirety.*

*For conditions of use, please consider all the information within.*

*This certificate may only be reproduced in its entirety and without change*

*To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

## Certificate Contents

Approved Site Application .....	2
Basis of Certification .....	2
Product Certified.....	3
Certified Performance .....	4
Description.....	19
General Notes .....	20

## Approved Site Application

*Any potential user should ensure, in consultation with the manufacturer, that the monitoring system is suitable for the intended application. For general guidance on monitoring techniques refer to the Environment Agency Monitoring Technical Guidance Notes available at [www.mcerts.net](http://www.mcerts.net)*

On the basis of the assessment and the ranges required for compliance with EU Directives this instrument is considered suitable for use on waste incineration and large coal-fired combustion plant applications. This CEM has been proven suitable for its measuring task (parameter and composition of the flue gas) by use of the QAL 1 procedure specified in EN14181, for LCPD/IED Chapter III and IED Chapter IV applications for the ranges specified. The lowest certified range for each determinand shall not be more than 1.5X the daily average emission limit value (ELV) for IED Chapter IV applications, and not more than 2.5X the ELV for IED Chapter III and other types of application.

## Basis of Certification

This certification is based on the following Test Report(s) and on Sira's assessment and ongoing surveillance of the product and the manufacturing process:

TÜV Rheinland Immissionsschutz und Energiesysteme GmbH test report No.: 936/21211670/B dated March 26th 2010


Sira Report 16A23053 Rev 1 dated 19<sup>th</sup> October 2010

Certificate No : Sira MC100182/04  
This Certificate issued : 21 December 2015

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

## Product Certified

The MAC GMS800 measuring system consists of the following parts:



1. Sample Probe	2. Heated Filter	3. Heated Sample Line	4. Gas Conditioning	5. Analyser
Model: M&C SP 2000	Model: N/A – Heated filter is integrated with M&C SP 2000 sample probe	Model: Not Stated Length: 50m	Model: MAK 10-2 / CSS-V2SK	Model: UNOR MULTOR DEFOR OXOR-P/E

Allowable variations could include:

- A different brand or model of sampling system of the same type, provided that there is evidence the alternative system works with similar types of CEM.
- Additional manifolds and heated valves used to allow more than one analyser to share a sampling system.

This certificate applies to all instruments fitted with software version T825\_090707\_1000 onwards.  
PC Software - Sopas ET 2.20 Build 2766 onwards, (serial number 10440002 onwards).

Certificate No : Sira MC100182/04  
This Certificate issued : 21 December 2015

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

## Certified Performance

The instrument was evaluated for use under the following conditions:

Ambient Temperature Range: +5 to +45°C with MAK 10-2 Cooler  
+10 to +40°C with CSS-V2SK Cooler

*If the instrument is supplied with an enclosure then the ambient temperature shall be monitored inside the enclosure to ensure that it stays within the above ambient temperature range.*

Instrument IP rating: IP 54  
Cooler IP rating: IP 34

Unless otherwise stated the evaluation was carried out on the following certification range: CO (UNOR) 0 to 75mg/m<sup>3</sup>, CO (MULTOR) 0 to 200 mg/m<sup>3</sup>, NO (UNOR) 0 to 100mg/m<sup>3</sup>, NO (MULTOR) 0 to 250 mg/m<sup>3</sup>, NO (DEFOR) 0 to 50 mg/m<sup>3</sup>, NO<sub>2</sub> (DEFOR) 0 to 50 mg/m<sup>3</sup>, NO<sub>x</sub> (UNOR) 0-100 mg/m<sup>3</sup>, SO<sub>2</sub> (UNOR) 0 to 75mg/m<sup>3</sup>, SO<sub>2</sub> (MULTOR) 0 to 250 mg/m<sup>3</sup>, SO<sub>2</sub> (DEFOR) 0 to 75 mg/m<sup>3</sup>, CH<sub>4</sub> (UNOR) 0 to 50 mg/m<sup>3</sup>, CH<sub>4</sub> (MULTOR) 0 to 286 mg/m<sup>3</sup>, N<sub>2</sub>O (UNOR) 0 to 50mg/m<sup>3</sup>, CO<sub>2</sub> (UNOR) 0 to 25%vol., CO<sub>2</sub> (MULTOR) 0 to 25%vol., O<sub>2</sub> (OXOR-P) 0 to 25%vol. and O<sub>2</sub> (OXOR-E) 0 to 25%vol..

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Response time						
CO (UNOR)					48s	<200s
CO (MULTOR)					137s	<200s
NO (UNOR)					44s	<200s
NO (MULTOR)					117s	<200s
NO (DEFOR)					33s	<200s
NO <sub>2</sub> (DEFOR)					61s	<200s
SO <sub>2</sub> (UNOR)					155s	<200s
SO <sub>2</sub> (MULTOR)					158s	<200s
SO <sub>2</sub> (DEFOR)					133s	<200s
CH <sub>4</sub> (UNOR)					49s	<200s
CH <sub>4</sub> (MULTOR)					71s	<200s
N <sub>2</sub> O (UNOR)					53s	<200s
CO <sub>2</sub> (UNOR)					44s	<200s
CO <sub>2</sub> (MULTOR)					95s	<200s
O <sub>2</sub> (OXOR-P)					33s	<200s
O <sub>2</sub> (OXOR-E)					44s	<200s
CO (UNOR) – (0 to 750mg/m <sup>3</sup> )					38s	<200s

Certificate No : Sira MC100182/04  
This Certificate issued : 21 December 2015

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
CO (UNOR) – (0 to 3000mg/m <sup>3</sup> )					39s	<200s
CO (MULTOR) – (0 to 2500mg/m <sup>3</sup> )					98s	<200s
NO (UNOR) – (0 to 1000mg/m <sup>3</sup> )					42s	<200s
NO (UNOR) – (0 to 2000mg/m <sup>3</sup> )					37s	<200s
NO (MULTOR) – (0 to 2500mg/m <sup>3</sup> )					100s	<200s
NO (DEFOR) – (0 to 1000mg/m <sup>3</sup> )					43s	<200s
NO (DEFOR) – (0 to 2000mg/m <sup>3</sup> )					37s	<200s
NO <sub>2</sub> (DEFOR) – (0 to 500mg/m <sup>3</sup> )					57s	<200s
SO <sub>2</sub> (UNOR) – (0 to 287mg/m <sup>3</sup> )					97s	<200s
SO <sub>2</sub> (UNOR) – (0 to 2000mg/m <sup>3</sup> )					43s	<200s
SO <sub>2</sub> (MULTOR) – (0 to 2000mg/m <sup>3</sup> )					96s	<200s
SO <sub>2</sub> (DEFOR) – (0 to 287mg/m <sup>3</sup> )					51s	<200s
SO <sub>2</sub> (DEFOR) – (0 to 2000mg/m <sup>3</sup> )					41s	<200s
CH <sub>4</sub> (UNOR) – (0 to 500mg/m <sup>3</sup> )					51s	<200s
CH <sub>4</sub> (MULTOR) – (0 to 500mg/m <sup>3</sup> )					65s	<200s
N <sub>2</sub> O (UNOR) – (0 to 500mg/m <sup>3</sup> )					40s	<200s
<i>Supplementary test 1 (NO<sub>x</sub> convertor):</i>					Note 3	
NO <sub>x</sub> (UNOR)					51s	<200s
CO (UNOR)					52s	<200s
<i>Supplementary test 2 (Cooler CSS-V2SK)</i>					Note 5	
NO (DEFOR)					63s	<200s
NO <sub>2</sub> (DEFOR)					74s	<200s
SO <sub>2</sub> (UNOR)					150s	<200s
SO <sub>2</sub> (DEFOR)					144s	<200s
CH <sub>4</sub> (UNOR)					41s	<200s
CH <sub>4</sub> (MULTOR)					83s	<200s
CO <sub>2</sub> (MULTOR)					81s	<200s
O <sub>2</sub> (OXOR-E)					46s	<200s

Certificate No : Sira MC100182/04  
This Certificate issued : 21 December 2015

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Repeatability standard deviation at zero point						
CO (UNOR)	0.30					<2.0%
CO (MULTOR)	0.10					<2.0%
NO (UNOR)	0.40					<2.0%
NO (MULTOR)		0.80				<2.0%
NO (DEFOR)	0.10					<2.0%
NO <sub>2</sub> (DEFOR)	0.00					<2.0%
SO <sub>2</sub> (UNOR)	0.20					<2.0%
SO <sub>2</sub> (MULTOR)		0.70				<2.0%
SO <sub>2</sub> (DEFOR)	0.20					<2.0%
CH <sub>4</sub> (UNOR)		0.60				<2.0%
CH <sub>4</sub> (MULTOR)	0.00					<2.0%
N <sub>2</sub> O (UNOR)	0.20					<2.0%
CO <sub>2</sub> (UNOR)	0.00					<2.0%
CO <sub>2</sub> (MULTOR)	0.00					<2.0%
O <sub>2</sub> (OXOR-P)	0.00					<0.2%
O <sub>2</sub> (OXOR-E)	0.20					<0.2%

Certificate No : Sira MC100182/04  
This Certificate issued : 21 December 2015

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Repeatability standard deviation at reference point						
CO (UNOR)	0.20					<2.0%
CO (MULTOR)	0.10					<2.0%
NO (UNOR)	0.20					<2.0%
NO (MULTOR)		0.60				<2.0%
NO (DEFOR)	0.20					<2.0%
NO <sub>2</sub> (DEFOR)			1.00			<2.0%
SO <sub>2</sub> (UNOR)	0.30					<2.0%
SO <sub>2</sub> (MULTOR)	0.20					<2.0%
SO <sub>2</sub> (DEFOR)	0.20					<2.0%
CH <sub>4</sub> (UNOR)			1.30			<2.0%
CH <sub>4</sub> (MULTOR)	0.20					<2.0%
N <sub>2</sub> O (UNOR)		0.50				<2.0%
CO <sub>2</sub> (UNOR)	0.20					<2.0%
CO <sub>2</sub> (MULTOR)	0.10					<2.0%
O <sub>2</sub> (OXOR-P)	0.20					<0.2%
O <sub>2</sub> (OXOR-E)	0.30					<0.2%
Lack of fit						
CO (UNOR)		0.67				<2.0%
CO (MULTOR)		1.00				<2.0%
NO (UNOR)	0.40					<2.0%
NO (MULTOR)		-0.80				<2.0%
NO (DEFOR)	0.40					<2.0%
NO <sub>2</sub> (DEFOR)		-0.80				<2.0%
SO <sub>2</sub> (UNOR)		0.95				<2.0%
SO <sub>2</sub> (MULTOR)			-1.88			<2.0%
SO <sub>2</sub> (DEFOR)		-0.93				<2.0%
CH <sub>4</sub> (UNOR)		0.80				<2.0%

Certificate No : Sira MC100182/04  
This Certificate issued : 21 December 2015

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
CH <sub>4</sub> (MULTOR)		-0.91				<2.0%
N <sub>2</sub> O (UNOR)		0.80				<2.0%
CO <sub>2</sub> (UNOR)			-1.00			<2.0%
CO <sub>2</sub> (MULTOR)			-1.64			<2.0%
O <sub>2</sub> (OXOR-P)	-0.07					<0.2%
O <sub>2</sub> (OXOR-E)	-0.10					<0.2%
CO (UNOR) – (0 to 750mg/m <sup>3</sup> )	0.45					<2.0%
CO (UNOR) – (0 to 3000mg/m <sup>3</sup> )		0.57				<2.0%
CO (MULTOR) – (0 to 2500mg/m <sup>3</sup> )		-1.00				<2.0%
NO (UNOR) – (0 to 1000mg/m <sup>3</sup> )	-0.25					<2.0%
NO (UNOR) – (0 to 2000mg/m <sup>3</sup> )	-0.25					<2.0%
NO (MULTOR) – (0 to 2500mg/m <sup>3</sup> )		-0.88				<2.0%
NO (DEFOR) – (0 to 1000mg/m <sup>3</sup> )	0.48					<2.0%
NO (DEFOR) – (0 to 2000mg/m <sup>3</sup> )		-0.65				<2.0%
NO <sub>2</sub> (DEFOR) – (0 to 500mg/m <sup>3</sup> )	0.34					<2.0%
SO <sub>2</sub> (UNOR) – (0 to 287mg/m <sup>3</sup> )		-0.98				<2.0%
SO <sub>2</sub> (UNOR) – (0 to 2000mg/m <sup>3</sup> )	0.47					<2.0%
SO <sub>2</sub> (MULTOR) – (0 to 2000mg/m <sup>3</sup> )		0.85				<2.0%
SO <sub>2</sub> (DEFOR) – (0 to 287mg/m <sup>3</sup> )		-0.98				<2.0%
SO <sub>2</sub> (DEFOR) – (0 to 2000mg/m <sup>3</sup> )	0.50					<2.0%
CH <sub>4</sub> (UNOR) – (0 to 500mg/m <sup>3</sup> )	-0.48					<2.0%
CH <sub>4</sub> (MULTOR) – (0 to 500mg/m <sup>3</sup> )		-0.70				<2.0%
N <sub>2</sub> O (UNOR) – (0 to 500mg/m <sup>3</sup> )	0.40					<2.0%
<i>Supplementary test 1 (NO<sub>x</sub> convertor):</i>						
NO <sub>x</sub> (UNOR)		-0.60				<2.0%
CO (UNOR)		0.93				<2.0%

Certificate No : Sira MC100182/04  
This Certificate issued : 21 December 2015

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*



Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
<i>Supplementary test 2 (Cooler CSS-V2SK)</i>						
NO (DEFOR)		0.80				<2.0%
NO <sub>2</sub> (DEFOR)		0.60				<2.0%
SO <sub>2</sub> (UNOR)		-1.00				<2.0%
SO <sub>2</sub> (DEFOR)		0.60				<2.0%
CH <sub>4</sub> (UNOR)	0.40					<2.0%
CH <sub>4</sub> (MULTOR)		0.80				<2.0%
CO <sub>2</sub> (MULTOR)		0.90				<2.0%
O <sub>2</sub> (OXOR-E)	0.09					<0.2%
Influence of ambient temperature zero and reference point						
CO (UNOR)			2.00			<5.0%
CO (MULTOR)				2.40		<5.0%
NO (UNOR)			2.00			<5.0%
NO (MULTOR)			1.90			<5.0%
NO (DEFOR)		0.60				<5.0%
NO <sub>2</sub> (DEFOR)			1.80			<5.0%
SO <sub>2</sub> (UNOR)				2.50		<5.0%
SO <sub>2</sub> (MULTOR)				-2.30		<5.0%
SO <sub>2</sub> (DEFOR)				2.40		<5.0%
CH <sub>4</sub> (UNOR)			1.60			<5.0%
CH <sub>4</sub> (MULTOR)			-1.90			<5.0%
N <sub>2</sub> O (UNOR)			-1.60			<5.0%
CO <sub>2</sub> (UNOR)				-2.40		<5.0%
CO <sub>2</sub> (MULTOR)		-0.80				<5.0%
O <sub>2</sub> (OXOR-P)	0.20					<0.5%
O <sub>2</sub> (OXOR-E)	0.24					<0.5%

Certificate No : Sira MC100182/04  
This Certificate issued : 21 December 2015

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Influence of sample gas pressure					N/A	<i>Test not applicable to extractive systems</i>
Influence of sample gas flow for extractive CEMS						
CO (UNOR)	0.30					<2.0%
CO (MULTOR)	0.11					<2.0%
NO (UNOR)	0.20					<2.0%
NO (MULTOR)	0.19					<2.0%
NO (DEFOR)		0.60				<2.0%
NO <sub>2</sub> (DEFOR)	0.40					<2.0%
SO <sub>2</sub> (UNOR)	0.40					<2.0%
SO <sub>2</sub> (MULTOR)	-0.28					<2.0%
SO <sub>2</sub> (DEFOR)	0.10					<2.0%
CH <sub>4</sub> (UNOR)		0.80				<2.0%
CH <sub>4</sub> (MULTOR)	0.30					<2.0%
N <sub>2</sub> O (UNOR)	0.18					<2.0%
CO <sub>2</sub> (UNOR)	0.12					<2.0%
CO <sub>2</sub> (MULTOR)	0.10					<2.0%
O <sub>2</sub> (OXOR-P)	0.04					<0.2%
O <sub>2</sub> (OXOR-E)	0.05					<0.2%
Influence of voltage variations 190 to 250V					No influence	<2.0% all gases
					No influence	<0.2% O <sub>2</sub>
Influence of vibration (10 to 60Hz (±0.3mm), 60 to 150Hz at 19.6m/s <sup>2</sup> )					N/A	<i>Test not applicable to extractive systems</i>

Certificate No : Sira MC100182/04  
This Certificate issued : 21 December 2015

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Cross-sensitivity at zero with interferents: O <sub>2</sub> , H <sub>2</sub> O, CO, CO <sub>2</sub> , CH <sub>4</sub> , N <sub>2</sub> O, NO, NO <sub>2</sub> , NH <sub>3</sub> , SO <sub>2</sub> , HCl						
CO (UNOR)				2.40		<4.0%
CO (MULTOR)		0.57				<4.0%
NO (UNOR)			1.56			<4.0%
NO (MULTOR)				3.58		<4.0%
NO (DEFOR)				3.72		<4.0%
NO <sub>2</sub> (DEFOR)				2.46		<4.0%
SO <sub>2</sub> (UNOR)				3.66		<4.0%
SO <sub>2</sub> (MULTOR)				3.85		<4.0%
SO <sub>2</sub> (DEFOR)			-1.08			<4.0%
CH <sub>4</sub> (UNOR)				-3.54		<4.0%
CH <sub>4</sub> (MULTOR)	<0.5				Note 2	<4.0%
N <sub>2</sub> O (UNOR)				-2.82		<4.0%
CO <sub>2</sub> (UNOR)			-1.88			<4.0%
CO <sub>2</sub> (MULTOR)	<0.5				Note 2	<4.0%
O <sub>2</sub> (OXOR-P)	<0.33				Note 2	<0.40%
O <sub>2</sub> (OXOR-E)	<0.33				Note 2	<0.40%
Supplementary test 1 (NO <sub>x</sub> convertor):					Note 3	
Interferent: NH <sub>3</sub>						
NO <sub>x</sub> (UNOR)	<0.5				Note 2	<4.0%
CO (UNOR)	<0.5				Note 2	<4.0%

Certificate No : Sira MC100182/04  
This Certificate issued : 21 December 2015

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
<i>Supplementary test 2 (Cooler CSS-V2SK)</i>					Note 5	
Interferent: H <sub>2</sub> O						
NO (DEFOR)	<0.5				Note 2	<4.0%
NO <sub>2</sub> (DEFOR)	<0.5				Note 2	<4.0%
SO <sub>2</sub> (UNOR)	<0.5				Note 2	<4.0%
SO <sub>2</sub> (DEFOR)	<0.5				Note 2	<4.0%
CH <sub>4</sub> (UNOR)	<0.5				Note 2	<4.0%
CH <sub>4</sub> (MULTOR)	<0.5				Note 2	<4.0%
CO <sub>2</sub> (MULTOR)	<0.5				Note 2	<4.0%
O <sub>2</sub> (OXOR-E)	<0.5				Note 2	<0.40%
Cross-sensitivity at reference with interferents: O <sub>2</sub> , H <sub>2</sub> O, CO, CO <sub>2</sub> , CH <sub>4</sub> , N <sub>2</sub> O, NO, NO <sub>2</sub> , NH <sub>3</sub> , SO <sub>2</sub> , HCl						
CO (UNOR)			1.43			<4.0%
CO (MULTOR)				3.38		<4.0%
NO (UNOR)				2.46		<4.0%
NO (MULTOR)			1.66			<4.0%
NO (DEFOR)				-3.40		<4.0%
NO <sub>2</sub> (DEFOR)				3.86		<4.0%
SO <sub>2</sub> (UNOR)				3.06		<4.0%
SO <sub>2</sub> (MULTOR)				3.81		<4.0%
SO <sub>2</sub> (DEFOR)				-3.88		<4.0%
CH <sub>4</sub> (UNOR)			-1.26			<4.0%
CH <sub>4</sub> (MULTOR)		-0.52				<4.0%
N <sub>2</sub> O (UNOR)				-2.38		<4.0%
CO <sub>2</sub> (UNOR)				-0.48		<4.0%
CO <sub>2</sub> (MULTOR)	0.40					<4.0%
O <sub>2</sub> (OXOR-P)	0.33					<0.40%
O <sub>2</sub> (OXOR-E)	0.33					<0.40%

Certificate No : Sira MC100182/04  
This Certificate issued : 21 December 2015

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
<i>Supplementary test 1 (NOx convertor):</i>					Note 3	
Interferent: NH <sub>3</sub>						
NO <sub>x</sub> (UNOR)	<0.5				Note 2	<4.0%
CO (UNOR)	<0.5				Note 2	<4.0%
<i>Supplementary test 2 (Cooler CSS-V2SK)</i>					Note 5	
Interferent: H <sub>2</sub> O						
NO (DEFOR)	<0.5				Note 2	<4.0%
NO <sub>2</sub> (DEFOR)	<0.5				Note 2	<4.0%
SO <sub>2</sub> (UNOR)	<0.5				Note 2	<4.0%
SO <sub>2</sub> (DEFOR)	<0.5				Note 2	<4.0%
CH <sub>4</sub> (UNOR)	<0.5				Note 2	<4.0%
CH <sub>4</sub> (MULTOR)	<0.5				Note 2	<4.0%
CO <sub>2</sub> (MULTOR)	<0.5				Note 2	<4.0%
O <sub>2</sub> (OXOR-E)	<0.33					<0.40%
Converter Efficiency					95.7%	>95%
					Note 4	

Certificate No : Sira MC100182/04  
This Certificate issued : 21 December 2015

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Measurement uncertainty						
CO (UNOR)					7.37%	7.5%
CO (MULTOR)					7.50%	7.5%
NO (UNOR)					11.71%	15%
NO (MULTOR)					11.90%	15%
NO (DEFOR)					11.06%	15%
NO <sub>2</sub> (DEFOR)					6.99%	15%
SO <sub>2</sub> (UNOR)					11.52%	15%
SO <sub>2</sub> (MULTOR)					10.73%	15%
SO <sub>2</sub> (DEFOR)					10.95%	15%
CH <sub>4</sub> (UNOR)					15.80%	22.5%
CH <sub>4</sub> (MULTOR)					10.36%	22.5%
N <sub>2</sub> O (UNOR)					5.71%	15%
CO <sub>2</sub> (UNOR)					5.20%	7.5%
CO <sub>2</sub> (MULTOR)					4.72%	7.5%
O <sub>2</sub> (OXOR-P)					2.36%	7.5%
O <sub>2</sub> (OXOR-E)					2.92%	7.5%
Field Trial						
Calibration function (field)						
CO (UNOR)					0.9003-0.9961	>0.90
CO (MULTOR)					0.9155-0.9945	>0.90
NO (UNOR)					0.9867-0.9990	>0.90
NO (MULTOR)					0.9568-0.9944	>0.90
NO (DEFOR)					0.9848-0.9983	>0.90
NO <sub>2</sub> (DEFOR)					0.9405-0.9969	>0.90
SO <sub>2</sub> (UNOR)					0.9621-0.9707	>0.90
SO <sub>2</sub> (MULTOR)					0.9549-0.9652	>0.90
SO <sub>2</sub> (DEFOR)					0.9453-0.9626	>0.90
CH <sub>4</sub> (UNOR)					0.9960-0.9999	>0.90

Certificate No : Sira MC100182/04  
This Certificate issued : 21 December 2015

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
CH <sub>4</sub> (MULTOR)					0.9958-0.9998	>0.90
N <sub>2</sub> O (UNOR)					0.9886-0.9995	>0.90
CO <sub>2</sub> (UNOR)					0.9846-0.9964	>0.90
CO <sub>2</sub> (MULTOR)					0.9467-0.9919	>0.90
O <sub>2</sub> (OXOR-P)					0.9764-0.9880	>0.90
O <sub>2</sub> (OXOR-E)					0.9019-0.9861	>0.90
<i>Supplementary test 1 (NO<sub>x</sub> convertor):</i>						
NO <sub>x</sub> (UNOR)					0.9940-0.9946	>0.90
Response time (field)					Maximum response time recorded was for  SO <sub>2</sub> 0-250 mg/m <sup>3</sup>  T <sub>90</sub> = 187 secs	<200s
Lack of fit (field)					Relative residuals do not exceed 2.0% (or 0.2% for O <sub>2</sub> ) of the certification range.	<2.0% <0.20% (O <sub>2</sub> )
Maintenance interval					Note 1	>8 days

Certificate No : Sira MC100182/04  
This Certificate issued : 21 December 2015

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Change in zero point over maintenance interval						
CO (UNOR)		0.80				<3.0%
CO (MULTOR)		0.80				<3.0%
NO (UNOR)				-2.10		<3.0%
NO (MULTOR)			1.90			<3.0%
NO (DEFOR)			1.30			<3.0%
NO <sub>2</sub> (DEFOR)				-2.40		<3.0%
SO <sub>2</sub> (UNOR)				-2.80		<3.0%
SO <sub>2</sub> (MULTOR)			1.50			<3.0%
SO <sub>2</sub> (DEFOR)			-1.40			<3.0%
CH <sub>4</sub> (UNOR)			1.80			<3.0%
CH <sub>4</sub> (MULTOR)	0.70					<3.0%
N <sub>2</sub> O (UNOR)			-1.10			<3.0%
CO <sub>2</sub> (UNOR)			-1.30			<3.0%
CO <sub>2</sub> (MULTOR)			-1.30			<3.0%
O <sub>2</sub> (OXOR-P)	-0.17					<0.20%
O <sub>2</sub> (OXOR-E)	-0.19					<0.20%
Change in reference point over maintenance interval						
CO (UNOR)			2.00			<3.0%
CO (MULTOR)				-2.60		<3.0%
NO (UNOR)				-2.40		<3.0%
NO (MULTOR)				2.90		<3.0%
NO (DEFOR)				2.40		<3.0%
NO <sub>2</sub> (DEFOR)				2.70		<3.0%
SO <sub>2</sub> (UNOR)			2.00			<3.0%
SO <sub>2</sub> (MULTOR)				-2.60		<3.0%
SO <sub>2</sub> (DEFOR)				-2.50		<3.0%
CH <sub>4</sub> (UNOR)				2.20		<3.0%

Certificate No : Sira MC100182/04  
This Certificate issued : 21 December 2015

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*



Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
CH <sub>4</sub> (MULTOR)			-1.80			<3.0%
N <sub>2</sub> O (UNOR)				2.70		<3.0%
CO <sub>2</sub> (UNOR)				2.40		<3.0%
CO <sub>2</sub> (MULTOR)				2.60		<3.0%
O <sub>2</sub> (OXOR-P)	0.19					<0.20%
O <sub>2</sub> (OXOR-E)	0.16					<0.20%
Availability					At least 98.9%	>95% (>98% for O <sub>2</sub> )
Reproducibility						
CO (UNOR)			2.00			<3.3%
CO (MULTOR)			1.60			<3.3%
NO (UNOR)				2.30		<3.3%
NO (MULTOR)			1.80			<3.3%
NO (DEFOR)				2.90		<3.3%
NO <sub>2</sub> (DEFOR)		1.00				<3.3%
SO <sub>2</sub> (UNOR)				3.20		<3.3%
SO <sub>2</sub> (MULTOR)			1.20			<3.3%
SO <sub>2</sub> (DEFOR)				3.20		<3.3%
CH <sub>4</sub> (UNOR)			1.90			<3.3%
CH <sub>4</sub> (MULTOR)	0.20					<3.3%
N <sub>2</sub> O (UNOR)			1.60			<3.3%
CO <sub>2</sub> (UNOR)			1.20			<3.3%
CO <sub>2</sub> (MULTOR)			1.30			<3.3%
O <sub>2</sub> (OXOR-P)	0.20					<0.20%
O <sub>2</sub> (OXOR-E)	0.20					<0.20%
Supplementary test 1 (NO <sub>x</sub> convertor):					Note 3	
NO <sub>x</sub> (UNOR)			1.80			<3.3%

Certificate No : Sira MC100182/04  
This Certificate issued : 21 December 2015

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

*Note 1* - CH<sub>4</sub> has a maintenance interval of 3 months, CO, NO, NO<sub>2</sub> and N<sub>2</sub>O have maintenance intervals of 4 weeks. SO<sub>2</sub>, CO<sub>2</sub> and O<sub>2</sub> have maintenance intervals of 2 weeks.

Automatic calibration of zero point shall be carried out at least once a week for all components besides O<sub>2</sub> (OXOR-P and OXOR-E) by using humidified ambient air. This procedure can be done automatically, controlled by the analyzer.

Automatic calibration of the reference point for all components shall be carried out weekly. Span gas or the calibration unit should be used for all components except oxygen. Span gas or humidified ambient air should be used to calibrate the oxygen channel (OXOR-P and OXOR-E).

*Note 2* - <0.5% of test gas concentration.

*Note 3* - After the field trial the system was modified by the manufacturer. A NO<sub>x</sub> converter was added to the modular measuring system. The modification was estimated as a Type 2 change of a suitably tested measuring system in accordance with EN 15267-3 which would invalidate the certification without additional tests. Therefore, the following supplementary tests were carried out:

- Linearity tests for the modules CO (UNOR), NO (UNOR)
- Response times for the modules CO (UNOR), NO (UNOR)
- 3 month field trial
- Monthly check of the converter efficiency
- AST inclusive comparison measurements with the SRM.

*Note 4* - No continuous decrease of convertor efficiency was observed.

*Note 5* - The measuring system may be operated with cooler type MAK10-2 by AGT Thermotechnik (original testing done on this model) as well as with cooler type CSS-V2SK by company M&C (Supplementary tests 2).

Certificate No : Sira MC100182/04  
This Certificate issued : 21 December 2015

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

## Description

The MAC GMS800 measuring system is a modular, multiple-component measuring system for continuous monitoring of flue gases. The sample gas is taken from the gas duct with the help of a sampling probe for gas and led to the measuring system via a heated sample gas tube. Subsequent analysis of the gas concentrations is carried out by gas analysing modules which can be individually adapted to the purposed application.

The tested measuring system comprised the following analyser modules:

- UNOR (CO, NO, SO<sub>2</sub>, CH<sub>4</sub>, N<sub>2</sub>O and CO<sub>2</sub>),
- MULTOR (CO, NO, SO<sub>2</sub>, CH<sub>4</sub> and CO<sub>2</sub>),
- DEFOR (NO, NO<sub>2</sub> and SO<sub>2</sub>),
- OXOR-P (O<sub>2</sub>)
- OXOR-E (O<sub>2</sub>).

### Additional measuring ranges:

Component	Module	Certification Range	Additional Ranges		Unit
			1	2	
CO	MAC GMS800 UNOR for CO	<b>0 - 75</b>	0-750	0-3000	mg/m <sup>3</sup>
CO	MAC GMS800 MULTOR for CO	<b>0-200</b>	0-2000	---	mg/m <sup>3</sup>
NO	MAC GMS800 UNOR for NO	<b>0-100</b>	0-1000	0-2000	mg/m <sup>3</sup>
NO	MAC GMS800 MULTOR for NO	<b>0-250</b>	0-2500	---	mg/m <sup>3</sup>
NO	MAC GMS800 DEFOR for NO	<b>0-50</b>	0-1000	0-2000	mg/m <sup>3</sup>
NO <sub>2</sub>	MAC GMS800 DEFOR for NO <sub>2</sub>	<b>0-50</b>	0-500	---	mg/m <sup>3</sup>
NO <sub>x</sub>	MAC GMS800 UNOR for NO <sub>x</sub>	<b>0-100</b>	0-1000	0-2000	mg/m <sup>3</sup>
SO <sub>2</sub>	MAC GMS800 UNOR for SO <sub>2</sub>	<b>0-75</b>	0-287	0-2000	mg/m <sup>3</sup>
SO <sub>2</sub>	MAC GMS800 MULTOR for SO <sub>2</sub>	<b>0-250</b>	0-2000	---	mg/m <sup>3</sup>
SO <sub>2</sub>	MAC GMS800 DEFOR for SO <sub>2</sub>	<b>0-75</b>	0-287	0-2000	mg/m <sup>3</sup>
CH <sub>4</sub>	MAC GMS800 UNOR for CH <sub>4</sub>	<b>0-50</b>	0-500	---	mg/m <sup>3</sup>
CH <sub>4</sub>	MAC GMS800 MULTOR for CH <sub>4</sub>	<b>0-286</b>	0-500	---	mg/m <sup>3</sup>
N <sub>2</sub> O	MAC GMS800 UNOR for N <sub>2</sub> O	<b>0-50</b>	0-500	---	mg/m <sup>3</sup>
CO <sub>2</sub>	MAC GMS800 UNOR for CO <sub>2</sub>	<b>0-25</b>	---	---	Vol.-%
CO <sub>2</sub>	MAC GMS800 MULTOR for CO <sub>2</sub>	<b>0-25</b>	---	---	Vol.-%
O <sub>2</sub>	MAC GMS800 OXOR-P for O <sub>2</sub>	<b>0-25</b>	---	---	Vol.-%
O <sub>2</sub>	MAC GMS800 OXOR-E for O <sub>2</sub>	<b>0-25</b>	---	---	Vol.-%

Certificate No : Sira MC100182/04  
This Certificate issued : 21 December 2015

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

## General Notes

1. This certificate is based upon the equipment tested. The Manufacturer is responsible for ensuring that on-going production complies with the standard(s) and performance criteria defined in this Certificate. The Manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management system shall be subject to regular surveillance according to 'Regulations Applicable to the Holders of Sira Certificates'. The design of the product certified is defined in the Sira Design Schedule V00 for certificate No. Sira MC100182/04
2. If certified product is found not to comply, Sira Certification Service should be notified immediately at the address shown on this certificate.
3. The Certification Marks that can be applied to the product or used in publicity material are defined in 'Regulations Applicable to the Holders of Sira Certificates'.
4. This document remains the property of Sira and shall be returned when requested by the company.

Certificate No : Sira MC100182/04  
This Certificate issued : 21 December 2015

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*